

Declaration of Conformity

File: F_Dec_MD_11

Page: 1 / 1

Date: September - 2020

Manufacturer : FEATHER SAFETY RAZOR CO., LTD.
3-70, Ohyodo-Minami 3-chome, Kita-ku, Osaka, 531-0075 JAPAN

Facility : FEATHER SAFETY RAZOR CO., LTD. MINO SITE
600-1, Matsumori, Mino-city, Gifu, 501-3753 JAPAN

Authorized Representative : pfm medical ag
Wankelstr. 60, 50996 Köln, GERMANY

We, FEATHER SAFETY RAZOR CO., LTD. located at 3-70, Ohyodo-Minami 3-chome, Kita-ku, Osaka, 531-0075 JAPAN herewith declare under our sole responsibility that the product(s) stated below are in conformity with the essential requirements and other relevant requirements of EC Directive 93/42/EEC. Any alterations made without our consent shall render this declaration null and void.

Product category : Blood Lancets

Product group : Blood Lancets

Classification : Class IIa (acc. to Rule 6 Annex IX of MDD 93/42/EEC)

Product name : **FEATHER BLOOD LANCET**

Model No. (Ref. No.) : 02.030.00.000

Intended purpose of use : Cutting instruments for blood collecting

Start of CE marking : Lot No. 11013670

GMDN-Code : Blood lancet, single-use (61579)

EC-Directive : Medical Device Directive 93/42/EEC


Conformity Assessment Route : Annex II excluding (4), Full quality assurance system

Standards applied : EN ISO 13485:2016/AC:2018, EN ISO 14971:2012,
EN ISO 10993-1:2009/AC2010, EN ISO 10993-5:2009, EN ISO 10993-10:2013,
EN ISO 10993-18:2009, EN ISO 11137-1:2015, EN ISO 11137-2:2015,
EN ISO 11737-1:2006/AC2009, EN ISO 11737-2:2009, EN ISO 11607-1:2009,
EN ISO 11607-2:2006, EN ISO 15223-1:2016, EN 1041:2008, EN 62366:2008,
EN ISO 14644-1:2015, JIS T 2107:2011, EN 62366:2008 AMENDMENT 1

Notified body : TÜV SÜD Product Service GmbH ID No.0123
Ridlerstraße 65, 80339 München, Germany

EC Certificate : No. G1 057582 0035 Rev.01

Place, Date : GIFU JAPAN, September 29, 2020

Signature :  Satoshi Mitsuishi

Title : Director